

PharmLabs San Diego Certificate of Analysis



Sample **YT0002**

Tested for **7ALKS**
 Sampled - Received **Jan 17, 2025** Reported **Jan 29, 2025**
 Analyses executed **RES, MIBIG, MICX, HME, KTM, SDR**

Laboratory note: COA Update: 1/29/25 - Photo & "Analyzed For" updated as per client request.

KTM - Kratom

Analyzed Jan 17, 2025 | Instrument HPLC VWD | Method SOP-KTM
 The expanded Uncertainty of the Kratom analysis is approximately ±7.81% at the 95% Confidence Level

Analyte	LOD ppm	LOQ ppm	Result %	Result mg/g	Result mg/Serving	Result mg/Unit	Sample photography
7-hydroxy Mitragynine (7HMG)	0.008	0.025	3.02	30.23	19.65	195.35	
Mitragynine (MITG)	0.018	0.054	ND	ND	ND	ND	
Speciogynine (SPEG)	0.007	0.02	ND	ND	ND	ND	
Speciociliatine (SPCL)	0.004	0.011	ND	ND	ND	ND	

HME - Heavy Metals

Analyzed Jan 20, 2025 | Instrument ICP/MSMS | Method SOP-005

Analyte	LOD ug/g	LOQ ug/g	Result ug/g	Limit ug/g
Arsenic (As)	0.0009	0.0027	0.03	1.5
Cadmium (Cd)	0.0005	0.0015	ND	0.5
Mercury (Hg)	0.0058	0.0174	ND	3
Lead (Pb)	0.0006	0.0018	0.02	0.5

MIBIG - Microbial

Analyzed Jan 21, 2025 | Instrument qPCR and/or Plating | Method SOP-007

Analyte	LOD CFU/g	LOQ CFU/g	Result CFU/g	Limit CFU/g
Shiga toxin-producing Escherichia Coli	1.0	1.0	ND	1
Salmonella spp.	1.0	1.0	ND	1
Aspergillus fumigatus	1.0	1.0	ND	1
Aspergillus flavus	1.0	1.0	ND	1
Aspergillus niger	1.0	1.0	ND	1
Aspergillus terreus	1.0	1.0	ND	1

RES - Residual Solvents

Analyzed Jan 20, 2025 | Instrument GC/FID with Headspace Analyzer | Method SOP-006

Analyte	LOD ug/g	LOQ ug/g	Result ug/g	Limit ug/g	Analyte	LOD ug/g	LOQ ug/g	Result ug/g	Limit ug/g
Propane (Prop)	0.044	0.4	ND	5000	Butane (But)	0.02	0.4	ND	5000
Methanol (Metha)	1.176	3.92	848.2	3000	Ethylene Oxide (EthOx)	0.08	0.4	ND	1
Pentane (Pen)	0.024	0.4	ND	5000	Ethanol (Ethan)	0.048	0.4	<LOQ	5000
Ethyl Ether (EthEt)	0.036	0.4	48.6	5000	Acetone (Acet)	0.044	0.4	66.1	5000
Isopropanol (2-Pro)	1.16	3.868	<LOQ	5000	Acetonitrile (Acetonit)	0.888	2.952	<LOQ	410
Methylene Chloride (MetCh)	0.04	0.4	ND	1	Hexane (Hex)	0.012	0.4	ND	290
Ethyl Acetate (EthAc)	0.032	0.4	176.3	5000	Chloroform (Clo)	0.028	0.4	ND	1
Benzene (Ben)	0.012	0.4	ND	1	1-2-Dichloroethane (12-Dich)	0.024	0.4	ND	1
Heptane (Hep)	0.012	0.4	<LOQ	5000	Trichloroethylene (TriClEth)	0.072	0.4	ND	1
Toluene	0.036	0.4	ND	890	Xylenes (Xyl)	0.012	0.4	ND	2170

MICx - Microbial X

Analyzed Jan 21, 2025 | Instrument Plating | Method SOP-007

Analyte	LOD CFU/G	LOQ CFU/G	Result CFU/G	Limit CFU/G
Total Yeast & Molds (TYM)	1.0	1.0	ND	
Listeria (LIS)	1.0	1.0	ND	
Gram Negative Bacteria (BTGN)	1.0	1.0	ND	
Total Viable Aerobic Bacteria (TVAB)	1.0	1.0	ND	

UI Unidentified
 ND Not Detected
 N/A Not Applicable
 NT Not Reported
 LOD Limit of Detection
 LOQ Limit of Quantification
 <LOQ Detected
 >ULOL Above upper limit of linearity
 CFU/g Colony Forming Units per 1 gram
 TNTC Too Numerous to Count



DCC license: **C8-0000098-LIC**
 DEA license: **RP0611043**
 ISO/IEC 17025:2017 Acc. 85368

Authorized Signature

Brandon Starr

Brandon Starr, Quality Assurance Manager
 Wed, 29 Jan 2025 13:45:02 -0800

